

APPENDIX 11 - Noise Baseline Report
Technical Details

APPENDIX 11-1 - Photographs of Monitoring Locations



Figure 1(a) Allahou-Bazi - Noise Monitoring Location



Figure 1(b) Allahou-Bazi – Blacksmiths Workshop



Figure 1(c) Allahou-Bazi – Mechanics Workshop



Figure 1(d) Allahou-Bazi – Church



Figure 2(a) Angovia - Noise Monitoring Location



Figure 2(b) Angovia – Activities near monitoring location



Figure 3 The Mine Site - Noise Monitoring Location



Figure 4(a) Akakro - Noise Monitoring Location



Figure 4(b) Akakro – Village Well



Figure 5 Allahou Port - Noise Monitoring Location & moto-taxi on road



Figure 6(a) Kouakougnanou - Noise Monitoring Location



Figure 6 (b) Kouakougnanou – Village water pump



Figure 7 (a) Kossou - Noise Monitoring Location



Figure 7 (b) Kossou – football pitch of Kossou 1 & 2 school group

APPENDIX 11-2 - Instrumentation Details

Kit 1 (2D Consulting Afrique)

Sound Level Meter Svan 595 Sound Level Meter
 Serial Number 15685
 Calibration Date ?


Calibrator: Svan 30A Calibrator
 Serial Number: 19389
 Calibration Date: 2 October 2014


Kit 2 (2D Consulting Afrique)


Sound Level Meter Cirrus Optimus Green Sound Level Meter
 Serial Number G056226
 Calibration Date: 10 February 2015


Calibrator: See above


APPENDIX 11-3 - Noise Monitoring Results – Time Histories

















Certificate of Calibration



Equipment Details

Instrument Manufacturer Cirrus Research plc
Instrument Type CR:515
Description Acoustic Calibrator
Serial Number 55759

Calibration Procedure

The acoustic calibrator detailed above has been calibrated to the published data as described in the operating manual. The procedures and techniques used to follow the recommendations of the IEC standard Electroacoustics – Sound Calibrators IEC 60942:2003, IEC 60942:1997, BS EN 60942:1998 and BS EN 60942:2003 where applicable. The calibrator's main output is 94.00 dB (1 Pa) and this was set within the 0.01 dB resolution of the test system, i.e. one hundredth of a decibel. Numbers in {parenthesis} refer to the paragraph in IEC 60942.

Calibration Traceability

The calibrator above was calibrated against the calibration laboratory standards held by Cirrus Research plc. These are traceable to International Standards {A.0.6}. The standards are:

Microphone Type	B&K4180	Serial Number	1893453	Calibration Ref.	S 6009
Pistonphone Type	B&K4220	Serial Number	613843	Calibration Ref.	S 5964

Calibration Climate Conditions

The climatic test conditions were all maintained within the permitted limits of IEC 60942:1997.

Temperature	{B.3.2}	Permitted band	15°C to 25°C
Humidity	{B.3.2}	Permitted band	30% to 90% RH
Static Pressure	{B.3.2}	Permitted band	85 kPa to 105 kPa
Ambient Noise Level	{B.3.3.6}	Max permitted level	64 dB(Z)

Measurement Results

The figures below are the Calibration Laboratory test limits for this model calibrator and have a smaller tolerance than those permitted in IEC 60942.

94 dB Output	94.01 dB	Permitted band	93.95 to 94.05dB
104 dB Output	dB	Permitted band	103.80 to 104.30dB
Frequency	1000 Hz	Permitted band	990 to 1010Hz

Uncertainty

With an uncertainty coefficient of $k=2$, i.e. a 95% confidence level, the uncertainty of each measure is

94 dB Output	± 0.13 dB	104 dB Output	± 0.14 dB
Frequency	± 0.1 Hz	Level Stability	± 0.04 dB

Calibrated by

Calibration Date

10 February 2014

Calibration Certificate Number

214947

This Calibration Certificate is valid for 12 months from the date above.

Cirrus Research plc, Acoustic House, Bridlington Road, Hunmanby, North Yorkshire, YO14 0PH

Telephone: +44 (0) 1723 891655 Fax: +44 (0) 1723 891742

Email: sales@cirrusresearch.co.uk

Certificate of Calibration



Equipment Details

Instrument Manufacturer Cirrus Research plc
Instrument Type CR:171B
Description Sound Level Meter
Serial Number G056226

Calibration Procedure

The instrument detailed above has been calibrated to the publish test and calibration data as detailed in the instrument hand book, using the techniques recommended in the latest revisions of the International Standards IEC 61672-1:2002, IEC 60651:1979, IEC 60804:2001, IEC 61260:1995, IEC 60942:1997, IEC 61252:1993, ANSI S1.4-1983, ANSI S1.11-1986 and ANSI S1.43-1997 where applicable.

Sound Level Meters: All Calibration procedures were carried out by substituting the microphone capsule with a suitable electrical signal, apart from the final acoustic calibration.

Calibration Traceability

The equipment detailed above was calibrated against the calibration laboratory standards held by Cirrus Research plc. These are traceable to International Standards {A.0.6}. The standards are:

Microphone Type	B&K4180	Serial Number	1893453	Calibration Ref.	S 6009
Pistonphone Type	B&K4220	Serial Number	613843	Calibration Ref.	S 5964

Calibrated by

A handwritten signature in black ink that reads 'T. A. Goodill'.

Calibration Date

10 February 2014

Calibration Certificate Number

214948

This Calibration Certificate is valid for 12 months from the date above.

Cirrus Research plc, Acoustic House, Bridlington Road, Hunmanby, North Yorkshire, YO14 0PH
Telephone: +44 (0) 1723 891655 Fax: +44 (0) 1723 891742
Email: sales@cirrusresearch.co.uk